

Major Tool & Machine, Inc.
1458 East 19th Street
Indianapolis, IN 46218-4289

MTM N/C: 19060

Page: 1
Date: 01/31/06
User ID: MCCORKLE

Customer: PRINCETON PLASMA PHYSICS LAB

Contact: LARRY SUTTON
E-Mail: S-04286-F

Telephone: 609-243-2441
Fax: 609-243-2021

Part: / SE122-049-1 PORT 4B FLANGE

Drawing ID: SE122-049 Revision: 1

Customer P.O.: S005243-F/Ln:7
Serial No: 7/9

Reported By: DOUG MCCORKLE
E-Mail: dMcCorkle@MajorTool.com

Telephone: 317-636-6433
Fax: 317-634-9420

Problem: PER B.P. MAG. PERM. TO BE LESS THAN 1.02
MATERIAL CHECKS GREATER THAN 1.05 BUT LESS THAN 1.1

Proposed Disposition:

CUSTOMER DISPOSITION REQUIRED

Number of additional pages: 0

Customer Disposition: ☐ Use As Is ☐ Rework ☐ Repair ☐ Scrap ☐ Replace

Technical Contact Approval: _____

Title: _____ **Date:** _____

Buyer Approval: _____

Title: _____ **Date:** _____

Major Tool Implemented By: _____

Title: _____ **Date:** _____

Nonconformance Report: Major Tool NC19060

This is for **SE122-049-1 PORT 4B FLANGE**

Problem:

PER B.P. MAG. PERM. TO BE LESS THAN 1.02

MATERIAL CHECKS GREATER THAN 1.05 BUT LESS THAN 1.1

Project Disposition:

Accept as is

Approvals:

Procurement Technical Representative

Art Brooks, Technical Assurance

Responsible Line Manager:

Project Quality Assurance: