
Customer: PRINCETON PLASMA PHYSICS LAB

Contact: Mike Viola
E-Mail: mviola@pppl.gov

Telephone: 609-243-3655
Fax: 609-243-3248

Part: SE120-004-47 / SEAL RETAINER SCREW

Drawing ID:

Revision:

Customer P.O.: S005243-F/Ln:7
Serial No./Qty:

Reported By: DOUG MCCORKLE
E-Mail: dMcCorkle@MajorTool.com

Telephone: 317-636-6433
Fax: 317-634-9420

Problem: seal retainer screw locked into nb cover after thermal cycling and vacuum testing. screw had no lubricant nor was it silver plated.
fab was able to weld onto head and freely remove all 8 screws without damage to threads.

Proposed Disposition:

Customer disposition required. The screws were destroyed during removal but the threaded holes are o.k. The same screws located in the port 12 and 4 flanges were removed without incident (they were not subjected to high temp).

Number of additional pages: 0

Customer Disposition: ☐ Use As Is ☐ Rework ☐ Repair ☐ Scrap ☐ Replace

Technical Contact Approval: _____

Title: _____ **Date:** _____

Buyer Approval: _____

Title: _____ **Date:** _____

Major Tool Implemented By: _____

Title: _____ **Date:** _____

Nonconformance Report: Major Tool NC19286

This is for **SE120-004-47 / SEAL RETAINER SCREW**

Problem:

seal retainer screw locked into nb cover after thermal cycling and vacuum testing. screw had no lubricant nor was it silver plated.

fab was able to weld onto head and freely remove all 8 screws without damage to threads.

The screws were destroyed during removal but the threaded holes are o.k.

The same screws located in the port 12 and 4 flanges were removed without incident (they were not subjected to high temp).

Doug McCorkle

Project Disposition:

Accept as is. PPPL will procure silver plated screws for future assemblies.

Approvals:

Procurement Technical Representative

Responsible Line Manager:

Project Quality Assurance: