

Nonconformance Reports:

Major Tool NCs:

17157, 17288, 17287, 17286, 17240, 17237, 17233, 17167, 17314, 17342

Project Disposition:

Substitute Trinos flanges per Contract change order #1 S5243 c-o 1 dated 6/22/2005.

Approvals:

Mike Viola

Digitally signed by Mike Viola
DN: CN = Mike Viola, C = US
Reason: I have reviewed this document
Date: 2005.08.22 13:41:12 -0400

Procurement Technical Representative

Brad Nelson

Digitally signed by Brad Nelson
DN: cn=Brad Nelson, c=US, o=CRRL, ou=FED,
email=bradn@crri.gov
Date: 2005.08.22 13:59:29 -0400

Responsible Line Manager:

Customer: PRINCETON PLASMA PHYSICS LAB

Contact: Mike Viola

E-Mail: mviola@pppl.gov

Telephone: 609-243-3655

Fax: 609-243-3248

Part: 401021 SPECIAL / DEL-SEAL CF HALF NIPPLE (SPECI

Customer P.O.: S005243-F

Drawing ID:

Revision:

Serial No./Qty: 30 Pieces

Reported By: DOUG MCCORKLE

E-Mail: dMcCorkle@MajorTool.com

Telephone: 317-636-6433

Fax: 317-634-9420

Problem: MAGNETIC PERMEABILITY TO BE 1.02 MAX
PARTS CHECK GREATER THAN 1.02
PARTS CHECK UP TO 1.05

Proposed Disposition:

CUSTOMER DISPOSITION REQUIRED. Propose – use as is.

Number of additional pages: _____

Customer Disposition: Use As Is Rework Repair Scrap Replace

Technical Contact Approval: _____

Title: _____

Date: _____

Buyer Approval: _____

Title: _____

Date: _____

Major Tool Implemented By: _____

Title: _____

Date: _____