

Status	2 - Disposition Needed	Trend	07-Out Of Tolerance
Department	NCSX	Division	WBS 142
Source/Org	Fabrication, Operations & Maintenance		
Item Dwg/Part#	A1 Mod. Coil	Procurement #	
RAP#	3234	Job Doc #	D-NCSX-MCF-004
Vendor			
RAP Title Modular Coil Fabrication - Post VPI Activities			

☐ HoldTag Applied**Nonconforming Condition (include requirement(s) violated):**

MCWF-A1; During the final Megger testing of the A1 modular coil an "arc over" or "short to ground" occurred at 5000 volts. D-NCSX-MCF-004 R3 paragraph 6.19.10 states an acceptance criteria for this test to be 7500 volts with >1K Meg ohms of insulation resistance.

Lot Size Recd	0	Sample Size Insp	0	<input type="checkbox"/> Lot Rejected	# Rejected	0
Reported By	Phelps C	Validated By	Boscoe J	Validated Date	11/08/07	

~~Disposition: Rework* __ Repair* __ Use As Is* __ Return To Vendor* __ Scrap* __~~

Please use p. 2 for disposition and approvals .

~~For rework or repair of vendor supplied equipments, fill in information below:~~

#Hours		\$Est Labor		\$G&A	
\$Material		\$Burden		\$Total	

Disposition By		Date	
Supervisor's Concur		Date	
Eng. Dept. Head Concur		Date	
WCO/Other		Date	

PQA/QC Mgr Dispos Concur		Date	
QC Field Verification By		Date	

Distribution

Cog J. Chrzanowski
Insp C. Phelps
Proj. Doc Control (when closed)
QC Files
Malsbury J
Boscoe J
T. Meighan
Simmons B
Williams M
Tyrrell M
Dudek L

Disposition: Rework____ Repair ____ Use As Is____ Return to Vendor____ Scrap____

For rework or repair of vendor supplied equipment, fill in information below:

Hours _____ \$ Est Labor _____ \$ G&A _____
\$ Material _____ \$ Burden _____ \$ Total _____

Disposition by _____

~~Supervisor's Concurrence~~ _____

Eng. Dept. Head Concurrence _____

Other (i.e., WCO/FPE) Concurrence _____

PQA/QC Mgr Disposition Concurrence _____

QA Field Verification by _____